

# 260/280 and 260/230 Ratios NanoDrop® ND-1000 and ND-8000 8-Sample Spectrophotometers

As absorbance measurements will measure any molecules absorbing at a specific wavelength, nucleic acid samples will require purification prior to measurement to ensure accurate results. Nucleotides, RNA, ssDNA, and dsDNA all will absorb at 260 nm and contribute to the total absorbance.

### 260/280

The ratio of absorbance at 260 nm and 280 nm is used to assess the purity of DNA and RNA. A ratio of ~1.8 is generally accepted as "pure" for DNA; a ratio of ~2.0 is generally accepted as "pure" for RNA. If the ratio is appreciably lower in either case, it may indicate the presence of protein, phenol or other contaminants that absorb strongly at or near 280 nm.

Some researchers encounter a consistent 260/280 ratio change when switching from a standard cuvette spectrophotometer to the NanoDrop<sup>®</sup> ND-1000 spectrophotometer. The three main explanations for this observation are listed below:

### CHANGE IN SAMPLE ACIDITY

Small changes in the pH of the solution will cause the 260/280 to vary\*. Acidic solutions will under-represent the 260/280 ratio by 0.2-0.3, while a basic solution will over-represent the ratio by 0.2-0.3. If comparing the ND-1000 or ND-8000 spectrophotometer to other spectrophotometers, it is important to ensure that the pH of an undiluted sample measured on the ND-1000 or ND-8000 is at the same pH and ionic strength as the diluted sample measured on the second spectrophotometer.

\* William W. Wilfinger, Karol Mackey, and Piotr Chomczynski, Effect of pH and Ionic Strength on the Spectrophotometric Assessment of Nucleic Acid Purity: *BioTechniques 22:474-481 (March 1997)* 

### WAVELENGTH ACCURACY OF THE SPECTROPHOTOMETERS

Although the absorbance of a nucleic acid at 260 nm is generally on a plateau, the absorbance curve at 280 nm is quite steeply sloped. A slight shift in wavelength accuracy will have a large effect on 260/280 ratios.

For example, a +/-1 nm shift in wavelength accuracy will result in a +/-0.2 change in the 260/280 ratio. Since many spectrophotometers claim a 1 nm accuracy specification, it is possible to see as much as a 0.4 difference in the 260/280 ratio when measuring the same nucleic acid sample on two spectrophotometers that are both within wavelength accuracy specification.

### NUCLEOTIDE MIX IN YOUR SAMPLE

The five nucleotides that comprise DNA and RNA exhibit widely varying 260/280 ratios\*\*. The following represent the 260/280 ratios estimated for each nucleotide if measured independently:

Guanine:	1.15
Adenine:	4.50
Cytosine:	1.51
Uracil:	4.00
Thymine:	1.47

The resultant 260:280 ratio for the nucleic acid being studied will be approximately equal to the weighted average of the 260/280 ratios for the four nucleotides present. It is important to note that the generally accepted ratios of 1.8 and 2.0 for DNA and RNA respectively, are "rules of thumb". The actual ratio will depend on the composition of the nucleic acid. Note: RNA will typically have a higher 260/280 ratio due to the higher ratio of Uracil compared to that of Thymine.

\*\* Leninger, A. L. Biochemistry, 2nd ed., Worth Publishers, New York, 1975

## 260/230

This ratio is used as a secondary measure of nucleic acid purity. The 260/230 values for "pure" nucleic acid are often higher than the respective 260/280 values. Expected 260/230 values are commonly in the range of 2.0-2.2. If the ratio is appreciably lower than expected, it may indicate the presence of contaminants which absorb at 230 nm.

Typical spectral pattern for Nucleic Acid (Figure 1)

atore		Print Screen Print Report	Recording Show Report	Measurement complete	User	2/16/2807 2:17 PM Default	E
13.30 -	Ove	riny control	Oeer graph each (	iomple 🔫		Sample Type	INA-58
12.00-		$ \land$				Sample ID	
8.00- 7.00- 6.00- 5.00-	$\bigvee$	/				Sam λ 230 At	plo # 4
4.00-	T					A-260 10 mm pc	
2.00 - 1.00 - 8.00 -			/			261	V280 1.8

Figure 1.

EDTA (Figure 2), carbohydrates and phenol all have absorbance near 230 nm. The TRIzol reagent is a phenolic solution which absorbs in the UV both at 230 nm and  $\sim$ 270 nm (Figure 3).




Figure 2.


Figure 4.

	Re-blank	Print Screen	n Recording	Make new BLANK		10/6/2006 2:38 PM	E
easure	Blank.	Print Report	Show Report	Measurement	User	Default	1000
12.221	Ovi	rlay control	Clear graph each	Sample 🔻		Sample RP	IA-40
11.00- 10.00- 9.00-							ol 1:100 Idution
8.00- 7.00- 6.00-	1	/	$\sim$			Samp	a # 🗌
5.00-	1	1				λ 260 Abs	
3.00-	/					A-288 18 mm patt	2.4
0.00-			_	-		250/	230 0.
220	250 24	210 200	270 200 290 Wavelength nm	oie oie oie oie	340 3		84.4



Guanidine HCL used for DNA isolations will absorb at  $\sim$ 230 nm (Figure 4) while guanidine isothiocyanate, used for RNA isolations will absorb at  $\sim$ 260 nm (Figure 5).

Aensure	Pro-blank Print Screen Blank Print Report		COLOR DE COL		User	3/11/2005 9:48 AM		
1.00-	Ovi	erlay control	Accumulate un	fil dear 👻		Sample DNA	-50	
0.90-							21-0.0M	
0.70-	X							
0.50-					-	Sample		
0.40-	1					1 A	0.682	
0.20-						A-280 10 mm path	-0.017	
0.10-	anne	Jan	No.				0 1.41	
-0.10-, 229	230 240	250 260 2	70 200 390 Wavelength nm	oce ate ate oor	340 D	ia	1.2	





Please contact Technical Support at 302-479-7707 or send an email to info@nanodrop.com for further information.